



MycAssay™ Aspergillus Serum

Platform	IAC Range	Positive Control Range	LoB	Software Version required
Roche LightCycler® 2.0	30.3 – 33.9	15.0 – 20.0	38.0	LightCycler® Software 4.1

A MycAssay™ Colour Compensation file should be applied to allow correct data analysis